

Human HSV2-IgG(herpes simplex virus 2-Immunoglobulin G) ELISA Kit

Catalogue No.: EH4957

Size: 96T

Reactivity: Human

Application: This immunoassay kit allows for the qualitative determination of HSV2-IgG in human serum.

Storage: 2-8°C.

Expiry Date: see kit label

Principle: Indirect

NOTE: FOR RESEARCH USE ONLY.

Kit Components

Item	Specifications(96T)	Storage
Micro ELISA Plate(Dismountable)	8 ×12	2-8°C /-20°C
HSV2-IgG Positive Control	0.5ml×1	2-8°C
HSV2-IgG Negative Control	0.5ml×1	2-8°C
Sample dilution buffer	12ml×1	2-8°C
HRP- Anti-Human IgG Antibody	12ml×1	2-8°C
TMB substrate A	6ml×1	2-8°C (Avoid Direct Light)
TMB substrate B	6ml×1	2-8°C (Avoid Direct Light)
Stop solution	6ml×1	2-8°C
Wash buffer (20X)	50ml×1	2-8°C
Plate Sealer	3pieces	
Product Description	1 copy	

Principle of the Assay

This kit was based on indirect enzyme-linked immune-sorbent assay technology. HSV2-Ag was pre-coated onto 96-well plates. The test samples were added to the wells, unbound conjugates were washed away with wash buffer. Then added HRP- Anti-Human IgG Antibody, if there were any HSV2-IgG in the samples, it would form a HSV2-Ag - HSV2-IgG - HRP- Anti-Human IgG complex. TMB substrates were used to visualize HRP enzymatic reaction. It was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The optical density of developed color is read with a suitable photometer at 450nm with a selected reference wavelength within 650 nm.

Precautions for Use

1. After opening and before using, keep plate dry.
2. Before using the Kit, balance the reagents at room temperature at least 30 mins.
3. Storage TMB reagents avoid light.
4. Washing process is very important, not fully wash easily cause a false positive.
5. Don't let Micro plate dry at the assay, for dry plate will inactivate active components on plate.
6. Don't reuse tips and tubes to avoid cross contamination.
7. Avoid using the reagents from different batches together.

Material Required But Not Supplied

1. Microplate reader (wavelength: 450nm)
2. 37°C incubator
3. Automated plate washer
4. Precision single and multi-channel pipette and disposable tips
5. Clean tubes and Eppendorf tubes
6. Deionized or distilled water

Manual Washing

Discard the solution in the plate without touching the side walls. Clap the plate on absorbent filter papers or other absorbent material. Fill each well completely with 350ul wash buffer and soak for 1 to 2 minutes, then aspirate contents from the plate, and clap the plate on absorbent filter papers or other absorbent material. Repeat this procedure two more times for a total of THREE washes.

Automated Washing

Aspirate all wells, then wash plate THREE times with 350ul wash buffer. After the final wash, invert plate, and clap the plate on absorbent filter papers or other absorbent material. It is recommended that the washer be set for a soaking time of 1 minute.

Sample Collection and Storage

Isolate the test samples soon after collecting, then, analyze immediately (within 2 hours). Or aliquot and store at -20°C for long term. Avoid multiple freeze-thaw cycles.

- Serum: Coagulate the serum at room temperature (about 1 hour). Centrifuge at approximately 1000 × g for 15 min. Analyze the serum immediately or aliquot and store at -20°C.

Note: Samples used within 5 days can be stored at 2-8°C; otherwise, they must be stored at -20°C or -80°C or liquid nitrogen to avoid loss of biological activity and contamination. Avoid multiple freeze-thaw cycles. Hemolytic samples are not suitable for this test.

Wash Buffer Preparation:

Dilute 50mL of Concentrated Wash Buffer into 950 mL of Wash Buffer with deionized or distilled water.

Assay Procedure

1. Wash plate 2 times before adding sample and control wells.
2. **Label** the sample wells, 3 Negative Controls, 1 Positive Controls and 1 blank well.
3. Add 100µL Negative Controls and Positive Controls to control wells.
4. Add 100µL sample dilution buffer to each sample well, then add 10µL sample to sample well and gently tap the plate to ensure thorough mixing. Seal the plate with a cover and incubate at 37°C for 30 min.
5. Remove the cover, and wash plate 5 times with Wash buffer and let the wash buffer stay in the wells for 1-2 minute each time.
6. Add 100 µL HRP- Anti-Human IgG Antibody to each well, except blank well.
7. Seal the plate with a cover and incubate at 37°C for 30 min.
8. Remove the cover, and wash plate 5 times with Wash buffer and let the wash buffer stay in the wells for 1-2 minute each time.
9. Add 50 µl of TMB substrate A and 50 µl of TMB substrate B into each well. Gently tap the plate to ensure thorough mixing. Cover the plate and incubate at 37°C in dark for 10 min. And the shades of blue can be seen in the Positive Controls. Negative Controls wells show no obvious color.
10. Add 50 µl of Stop solution into each well and mix thoroughly. The color changes into yellow immediately.

11. Read the O.D. absorbance at 450 nm in a microplate reader immediately after adding the stop solution.
(Use the blank well to set zero)

Data Analysis

NCx: Mean Absorbance of Negative Control

Cutoff Value = **NCx**+0.10

Quality control

NCx: $A \leq 0.1$ (If $A \leq 0.05$, set NCx as 0.05. If $0.05 < A \leq 0.1$, set NCx to the actual value. If $A > 0.1$, it should be retested)

Blank well: $A \leq 0.08$

Positive control: $A \geq 0.3$

Note: Sample with absorbance values $<$ Cutoff Value are NON-REACTIVE and are considered NEGATIVE for HSV2-IgG.

Sample with absorbance values \geq Cutoff Value are considered POSITIVE for HSV2-IgG.